



DECLARATION OF CONFORMITY

Ultradent Products Inc. has evaluated the following product by using the Conformity Assessment Procedure of Annex VII of the Medical Device Directive 93/42/EEC, as amended by 2007/47/EEC:

VALO Cordless (Including battery pack, battery recharger, power supplies)

and confirms that the product is compliant with the Essential Requirements of Annex I of the Medical Device Directive 93/42/EEC. Technical documentation is located in the Regulatory Affairs Department.

This product system is classified as Class I medical device according to the Medical Device Directive 93/42/EEC, Section III Classification 3.3 Rule 12.

UMDNS Code: 16386: Lights, Dental Activator

GMDN Code: 35775: Light Dental, Polymerization Activator

Ultradent Products GmbH
Am Westhover Berg 30
51149 Cologne
Germany

Marie Hess
Marie Hess
Sr. Regulatory Affairs Associate


5-Mar-15
Date

State of Utah
County of Salt Lake

Subscribed and sworn to before me on this 5 day of March 2015

By Marie Hess

Diane E. Rogers
Diane Rogers, Notary Public



This document is in force as long as the following EC certificates are valid:

EC Certificate 44 232 090234
EN ISO 13485 44 221 090234